

FOR VETERINARY USE ONLY



ZOETIS DIAGNOSTICS

vetscan OptiCell

User Manual

2024

Disclaimer

For Veterinary Use Only.

No part of this publication may be reproduced or transmitted in any form or by any means, electronic or mechanical, including photocopy, recording, or any information storage or retrieval system, without permission in writing from Zoetis. Zoetis reserves the right to modify its products and services at any time. This user's manual is subject to change without notice. Although prepared to ensure accuracy, Zoetis assumes no liability for errors, or for any damages resulting from application or use of this information. No rights are granted for diagnostic uses. No rights are granted for use of the product for commercial application of any kind, including but not limited to manufacturing, quality control, or commercial services, such as contract services or fee for services. It is the responsibility of the purchase/end user to acquire any additional intellectual property rights that may be required.

The analyser's protection will be impaired if used in a manner not specified by the manufacturer. Contact your Zoetis representative for details of warranty coverage, exclusionary defects, and warranty period. If any defects occur in the analyser during active warranty period, Zoetis will repair or replace the defective parts free of charge. The analyser's protection will be impaired if used in a manner not specified by the manufacturer. Contact your Zoetis representative for details of warranty coverage, exclusionary defects, and warranty period.



Caution: Animal blood is a potential source of zoonotic diseases. We recommend wearing personal protective equipment when handling animal blood or devices used for measuring animal blood. Be sure to follow local occupational health and safety regulations.

Read all instructions before using the Vetscan OptiCell™

The information in this manual is relevant to Vetscan OptiCell Model Number OptiCell 300.

EU Declaration of Conformity

Manufacturer

Zoetis, Inc. 333 Portage
Street Kalamazoo, MI
49007, USA

The Directives covered by this Declaration

2014/35/EU Low Voltage Directive 2014/30/EU
Electromagnetic Compliance Directive 2011/65/EU
RoHS Directive including amendments

The current EU Declaration of Conformity is on file at Zoetis. The end-user license agreement and Zoetis privacy policy may be retrieved through the Vetscan Hub user interface.

Table of Contents

1	General Information	6
1.1	Introduction and Intended Use	6
2	Set-Up, Overview, and Operating Instructions	7
2.1	Unpacking	7
2.2	Installation Guidelines	8
2.3	User Information and Precautions	10
2.4	Main Components	11
	2.4.1 Ventilation	12
2.5	Power States	13
2.6	Additional components (not included)	13
	2.6.1 Vetscan Hub and Router	13
2.7	Basic Operating Instructions	14
	2.7.1 Vetscan OptiCell Haematology Cartridge and Vetscan OptiCell Haematology Blood Samplers	14
	2.7.2 Running a Sample	15
	2.7.3 Adding Blood into the OptiCell Haematology Blood Samplers Via Water-Repellent Sheet	16
	2.7.4 Add OptiCell Haematology Blood Sampler into OptiCell Haematology Cartridge	16
3	Technical Specifications	17
4	Maintenance	19
4.1	Monitoring System and Quality Control	19
	4.1.1 Internal Monitoring System	19
	4.1.2 Quality Control Cartridge	20
4.2	Changing the Fan Filter	21
4.3	Software Updates	22
4.4	Service	22
4.5	Disposal of the Vetscan OptiCell	22
5	Technical Support and Troubleshooting	23
5.1	Technical Support	23
5.2	Errors and Warning Flags Displayed on Vetscan Hub	23
5.3	Preanalytical Guidelines	24
5.4	Error message displayed on Vetscan Hub	25

Table of Contents

6	Performance Specifications	31
6.1	Vetscan OptiCell Reportable Ranges (Linearity)	31
7	Certifications	32
7.1	Countries of compliance	32
7.2	USA Federal Communications Commission (FCC)	32
7.3	Industry Canada (IC) Warnings	32
7.4	Warranty	32
8	Contact information	33
8.1	Manufacturer	33
8.2	EU importer	33
8.3	AU importer	33
8.4	Technical support	33
9	Symbols	34

1. General Information

1.1 Introduction and intended use

Vetscan OptiCell is a fully automated point-of-care diagnostic haematology analyser, featuring a combination of three technological innovations to provide consistent and accurate results:¹ microfluidic Viscoelastic Focusing (VEF), Imaging-based AI algorithms, and a self-contained single-use cartridge system.

Innovative Artificial Intelligence (AI) driven cellular classification technology facilitates an automated complete blood count (CBC) analysis of 22 parameters using venous whole blood samples (K2/K3 EDTA anticoagulated) in validated species via a cartridge-based system.

✓ White blood cell (WBC) count	✓ Red blood cell (RBC) count
✓ Neutrophil percentage	✓ Haemoglobin (HGB)
✓ Neutrophil absolute count	✓ Haematocrit (HCT)
✓ Lymphocyte percentage	✓ Mean cell volume (MCV)
✓ Lymphocyte absolute count	✓ Mean corpuscular haemoglobin (MCH)
✓ Monocyte percentage	✓ Mean corpuscular haemoglobin concentration (MCHC)
✓ Monocyte absolute count	✓ Red blood cell distribution width (RDW)
✓ Eosinophil percentage	✓ Reticulocyte count
✓ Eosinophil absolute count	✓ Reticulocyte percentage
✓ Basophil percentage	✓ Platelet count (PLT)
✓ Basophil absolute count	✓ Mean platelet volume (MPV)

Vetscan OptiCell provides highly accurate¹ automated haematological cellular analysis with a simplified sample preparation workflow, minimal analyser maintenance, and no calibration requirement. The Vetscan OptiCell offers user guidance and sample alerts, via the Vetscan Hub™, when cellular morphological abnormalities or findings are detected. The Vetscan OptiCell analyser and Vetscan OptiCell Haematology Cartridges are intended for veterinary and professional use only.

Reference: 1. Data on file. Study Report No. DHXMZ-US-24-235, 2024, Zoetis Inc.



If the Vetscan OptiCell is used in any way other than described in this manual, the analyser may not operate as intended, may produce inaccurate or no results, and may pose a safety hazard.

2. Set-Up, Overview, and Operating Instructions

2.1 Unpacking

When unpacking the analyser, follow these guidelines:

- Remove Vetscan OptiCell from the shipping carton.
- Place the analyser on a clean, level surface free of animal hair, dust, and other contaminants.
- Do not stack anything on top of Vetscan OptiCell analyser.
- Do not place in direct sunlight, near an air-conditioner or heat source as this may increase the temperature outside of the specified range.

Check the components received with Vetscan OptiCell against the list and Figure 1 below to make sure all items required to set up the analyser are included.

- A. Vetscan OptiCell Analyser
- B. Power cable
- C. Power supply
- D. Ethernet cable
- E. Spare Fan Filter
- F. Quick Start Guides




Figure 1: Vetscan OptiCell components

2. Set-Up, Overview, and Operating Instructions

2.2 Installation Guidelines

Before setting up Vetscan OptiCell:

- Confirm room temperature is between 17 °C to 27 °C. If the temperature increases outside the specified operating range the system will display a warning message on the Vetscan Hub.
- Confirm the room humidity is between 10% and 90% relative humidity, non-condensing.
- Place the Vetscan OptiCell analyser on a clean, flat, and stable surface not in direct sunlight.
- Do not stack anything on top of the Vetscan OptiCell analyser.
- Ensure there is proper room ventilation and at least 10 cm (3.9 in) from the wall for access to the power connection and Ethernet port.

2. Set-Up, Overview, and Operating Instructions

Connecting the Vetscan OptiCell:

- Attach the AC power cord (with electrical wall plug) to the power supply.
- Connect the Vetscan OptiCell to the power supply.
- Plug the AC power cord into a grounded electrical outlet powered directly from AC mains infrastructure.
- Power supply must be operated at an ambient temperature, not to exceed 40° C maximum operating temperature.
- Make sure all connections are secure.



To prevent power surges or power drain, do not plug the analyser into the same circuit as a centrifuge or any other high-current device. If this is not possible, use an ancillary power conditioner or UPS.

- Connect one side of the supplied Ethernet cable to the Vetscan OptiCell analyser.
- Connect the other side of the Ethernet cable to the same network as the Vetscan Hub (via a router, switch, or direct Ethernet port access)
- The network does not need to be connected to the internet to operate the Vetscan OptiCell. However, active internet connection is needed to send results to the Practice Information Management Software (PIMS) and get full access to online Vetscan Hub application features.




Figure 2: Vetscan OptiCell Quick Installation Guide

2. Set-Up, Overview, and Operating Instructions

2.3 User Information and Precautions

Read the instructions in this User's Manual carefully before using the Vetscan OptiCell. Follow all instructions, warnings and precautions and keep the manual available for future reference. The Vetscan OptiCell analyser and Vetscan OptiCell Haematology consumable tests (cartridges and blood samplers) are intended for veterinary use only.



If the Vetscan OptiCell is used in any way other than described in this manual, the analyser may not operate as intended, may produce inaccurate or no results, and may pose a safety hazard.

- Always follow local regulations.
- Attempting to open or dismantle the analyser may cause electric shock and will void the warranty.
- Do not place objects on top of the analyser.
- Protect the analyser from liquids including exposure to wet locations.
- Use only the power supply provided. Use of any other power supply may damage the analyser, thus voiding the warranty. Usage of non-authorized power supplies may also be a safety hazard.
- The power supply must be connected to a main power outlet that is grounded.
- The Power Supply must be operated at an ambient that is limited to 40°C max and powered directly from AC mains infrastructure that provide 100-240VAC
- Ensure the power cord has a minimum rating of 10 A
- Never use a damaged power cord or power supply. Replace power cord if damaged.
- The analyser and test kits are operational within ideal temperature range of 17-27 °C, 10-90% RH, non-condensing.
- Never expose the Vetscan OptiCell to a heat source or to direct sunlight.
- When stored in a cold or warm area, let the analyser acclimate to 17-27 °C in advance of powering on the analyser.
- Do not place the analyser near centrifuges due to vibration and/or electrical interferences.
- It is recommended to inspect Vetscan OptiCell analyser for damage. If damaged, immediately contact Zoetis Diagnostic Technical Support. Contact information can be found at <https://www.zoetisdiagnostics.com.au>.
- If the Vetscan OptiCell analyser is accidentally dropped, immediately contact Zoetis Diagnostic Technical Support. Contact information can be found at <https://www.zoetisdiagnostics.com.au>.
- Incompatible, used, or expired cartridges are not valid and will result in an error message.
- Before inserting the test, make sure that there are no foreign objects in the sample drawer. Vetscan OptiCell Haematology Cartridge should slide easily into the drawer with a gentle push, after the sample run has been initiated via Vetscan Hub.

2.4 Main Components

Front

- Anti-stack lid
- LED light band
- Sample door

Back

- Air inlet and outlet ventilation
- Power button
- Power input
- Fan filter
- LAN connector




Figure 3: Vetscan OptiCell

2. Set-Up, Overview, and Operating Instructions

2.4.1 Ventilation




Figure 4: Vetscan OptiCell ventilation

 **The air inlet is located on the lower back panel of the analyser and contains a removable fan filter. To remove the fan filter, see the maintenance section.**

Reader Status	Light Colour
Start up/initialising/identify device	Breathing white
Idle	Steady white
Ready for operator to insert OptiCell Haematology Cartridge	Pulsing green
Analysing	Steady green
Analysis complete	Steady white
Maintenance procedure (software update, quality control cartridge)	Steady blue
Low power mode	After 1 hour of inactivity, LED turns off (analyser remains powered on and connected to Vetscan Hub).
Error	Breathing red. If error is not resolved after 5 minutes, the red LED will remain steady until addressed.

2.5 Power States

To power on, simply press the power button located on the lower back panel on Vetscan OptiCell. Vetscan OptiCell will initialise and connect with the Vetscan Hub ecosystem. The analyser will automatically enter idle mode when not in use. The primary method for exiting idle mode of Vetscan OptiCell is through Vetscan Hub.

2.6 Additional Components (not included)

- Vetscan Hub
- A router




Figure 5: Vetscan Hub and router (not included).

2.6.1 Vetscan Hub

Vetscan OptiCell is controlled from a user interface hosted on Vetscan Hub (not included). Please refer to the Vetscan Hub User Manual: Operating System for information related to the operation of the Vetscan Hub. The Vetscan Hub and router (see Figure 5) are not part of Vetscan OptiCell packaging.

2. Set-Up, Overview, and Operating Instructions

2.7 Basic Operating Instructions

2.7.1 Vetscan OptiCell Haematology Cartridge and Vetscan OptiCell Haematology Blood Samplers

Vetscan OptiCell introduces innovative technology using a combination of AI-powered machine vision (digital) imaging with flow cytometry that provides highly accurate and reliable complete blood cell count (CBC) results.¹ The blood sample and reagents are contained within the OptiCell Haematology Cartridge design, and flow through the reagent chambers where they react with test-specific reagents in preparation for analysis. During the analysis, blood cells are focused, aligned, and analysed within the OptiCell Haematology Cartridge's proprietary design. Cell populations are differentiated by AI-powered machine learning algorithms which extract hundreds of cellular features, including cell size, morphology, and staining properties. Test results are presented via the Vetscan Hub display.

Combining the Vetscan OptiCell Haematology Cartridge and Vetscan OptiCell Haematology Blood Sampler together forms a self-contained, disposable, consumable unit used for the collection and preparation of a single blood sample for an automated complete blood count haematology analysis.

Figure 6 depicts an overview of Vetscan OptiCell Haematology Cartridges and Vetscan OptiCell Haematology Blood Samplers. Vetscan OptiCell haematology testing requires one OptiCell Haematology Cartridge and one OptiCell Haematology Blood Sampler per test. The blood sample is added into the OptiCell Haematology Blood Sampler capillary tubes, prior to inserting in the OptiCell Haematology Cartridge.

Reference: 1. Data on file. Study Report No. DHXMZ-US-24-235, 2024, Zoetis Inc.




Figure 6: Generic view of Vetscan OptiCell Haematology Cartridge and Blood Sampler

2. Set-Up, Overview, and Operating Instructions

2.7.2 Running a Sample

Testing workflow begins with placing a test order via Practice Information Management Software (PIMS) or directly through the Vetscan Hub screen. To order a new test, enter in patient-specific data and select Haematology – OptiCell.

Once the test order is submitted, select the pending order under the test menu and press ‘start test’. The LED light band will begin to pulse green when OptiCell analyser is ready to accept the OptiCell Haematology Cartridge and begin the analysis.

***Note - The sample door remains locked until the LED light band switches to pulsing green.**

Next, follow the on-screen guidance to load the EDTA whole blood sample into the OptiCell Haematology Blood Sampler capillary tubes. The OptiCell Haematology Blood Sampler must be fully snapped into the OptiCell Haematology Cartridge before initiating a test run. After the OptiCell Haematology Blood Sampler is snapped into a OptiCell Haematology Cartridge, immediately insert the test into the analyser sample door.

2.7.3 Adding Blood into the OptiCell Haematology Blood Samplers Via Water-Repellent Sheet

1. Gently invert blood sample tube 15-20 times immediately before use
2. Open the EDTA anticoagulated tube cap and use a pipette to draw out blood from sample tube.
3. Gently transfer 1-2 drops of blood on clean sample sheet with a hydrophobic surface (such as a water-repellent sheet)
4. Fill both OptiCell Haematology Blood Sampler capillaries with blood directly from the hydrophobic surface and ensure capillaries are fully filled, with no air bubbles present.
5. If air bubbles are detected in the capillary tubes, discard, and use a new sampler.

***Note - It is not necessary to wipe off any excess blood on the capillary tubes.**




Figure 7: Adding Blood into the OptiCell Haematology Blood Samplers Via Water-Repellent Sheet

2. Set-Up, Overview, and Operating Instructions

2.7.4 Add OptiCell Haematology Blood Sampler into OptiCell Haematology Cartridge

1. Confirm the correct patient and test has been selected according to the Vetscan Hub display.
2. Following the on-screen guidance of the Vetscan Hub display, snap the loaded Vetscan OptiCell Haematology Blood Sampler into the Vetscan OptiCell Haematology Cartridge by applying gentle and steady pressure until it fully snaps into place.*

***Note - Do not touch the microfluidic chip area or apply pressure onto any of the reagent wells.**

- Sampler must be fully snapped in place, with the flat end of the sampler aligned with a flush edge of the cartridge.*

***Note - Listen for a 'click' sound to confirm it is fully snapped into place.**

- By design, the sampler cannot be removed from the cartridge once it is secured in place. Do not attempt to remove the blood sampler from the cartridge.

3. Insert OptiCell Haematology Cartridge into the analyser sample door once the LED light band begins to pulse green. The sample door remains closed and locked until the LED light band begins pulsing green.*

***Note - After initiating the haematology test run via Vetscan Hub, the user has a 10-minute timeout window to insert the cartridge into the analyser sample door.**

- OptiCell Haematology Cartridge can only be entered into the analyser with the correct orientation.
- Ensure the haematology cartridge microfluidic chip area is not touched or smudged with debris, as this may result in a test error.

In < 6 minutes, the haematology test results will display on the Vetscan Hub screen for interpretation.

Reference: 1. TBD study on file.




Figure 8: Add OptiCell Haematology Blood Sampler into OptiCell Haematology Cartridge

3. Technical Specifications

Table 1: Technical Specifications

Parameter	Specification
Assay	Complete Blood Count (CBC) + 5-part WBC Differential
Technology	Flow Cytometry, Microfluidic Viscoelastic Focusing (VEF) , and Digital Imaging with Artificial Intelligence.
Cellular Focusing Technology	Viscoelastic Focusing (VEF)
Sample Type	K2/K3 EDTA anticoagulated venous whole blood (canine or feline)
Sample Preparation	Automated sample and reagent preparation within a single-use, self-contained Vetscan OptiCell Haematology Cartridge
Sample Volume	40 µL total (2 capillary tubes x 20 µL each)
Test duration	< 6 minutes
Analyser Environmental Conditions	Operating environment of ambient room temperature, range of 17 °C to 27 °C.
Consumable Test Storage and Environmental Conditions	Storage and operating environmental conditions: <ul style="list-style-type: none"> • Ambient temperature range of 17 °C to 27 °C • Transportation Conditions: • Temperature range of 0°C to 60°C. • Relative humidity range of 10% to 90%, non-condensing humidity.
Operating Voltage	19 V DC, 6.3 A
Power Supply	100 – 240 VAC, 4 – 2 A, 50/60 Hz
Power Consumption	120 W
Dimensions (L x W x H)	Height 317.5 mm x Width 198 mm x Depth 249 mm
Weight	8 kg
Calibration Requirements	The OptiCell is factory-calibrated for optimal performance. There is no manual calibration requirement.

3. Technical Specifications

Parameter	Specification
Maintenance Requirements	Minimal maintenance requirement is a periodic check of the fan filter (located on the analyser back panel) for any fur or debris build up.
Service and Repair	Contact Zoetis Diagnostic Technical Support.
OptiCell Haematology Cartridge and OptiCell Haematology Blood Sampler Composition	Propylene Glycol, Sodium Chloride, Dextran. For veterinary use only! Not a hazardous material.

Reference: 1. Data on file. Study Report No. DHXMZ-US-24-235, 2024, Zoetis Inc.

4. Maintenance

Vetscan OptiCell requires minimal user maintenance. A fan filter cover is located on the analyser back panel. Check for any debris or excessive fur buildup on the fan filter area every few months, depending on the laboratory environment. Replace the fan filter as needed, and do not attempt to wash/reuse an existing filter. Proper maintenance of the analyser will assure reliable operation. A visual inspection of the analyser for damage is recommended. Clean the exterior of Vetscan OptiCell as needed with a soft, damp cloth and 70% isopropyl alcohol. Follow universal precautions when removing spills on or near the Vetscan OptiCell analyser.



Do not use Vetscan OptiCell if it appears to be damaged. Disconnect the power and contact Zoetis Diagnostic Technical Support.

Do not wash or reuse fan filters. Doing so may compromise the integrity of the fan filter material.

4.1 Monitoring System and Quality Control

Unlike traditional haematology devices, the Vetscan OptiCell analyser includes a robust internal monitoring system that eliminates the need for routine external quality controls. If external quality control is needed, it is recommended to use the OptiCell Quality Control Cartridge (QCC) once per month (every 30 days) to confirm optimal analyser performance. Under certain troubleshooting circumstances, Zoetis Diagnostic Technical Support may also recommend the use of the OptiCell QCC.

4.1.1 Internal Monitoring System

The OptiCell analyser employs a comprehensive set of internal system evaluations during normal operation to ensure that all components are performing as expected. The monitoring system tests specific parameters which correspond to each step of the sample analysis. This includes preanalytical sample integrity checks, evaluation of Vetscan OptiCell Haematology Cartridge components (including lot and expiry date), and vigorous internal checks for system optical integrity and stability. The Vetscan OptiCell Haematology Cartridge and reagents are checked for the presence of contamination and/or air bubbles.¹

Reference: 1. Data on file. Study Report No. DHXMZ-US-24-235, 2024, Zoetis Inc.

4. Maintenance

4.1.2 Quality Control Cartridge

A multi-use Quality Control Cartridge (QCC) is available to complement the comprehensive internal monitoring system built within the Vetscan OptiCell analyser for performance verification. The specialised OptiCell QCC measurement area contains fixed cell-like targets for analysis, which eliminates the need for traditional liquid QC materials. No additional reagents or external liquid control material is needed to run the OptiCell QCC. The OptiCell QCC is stored at ambient room temperature.²

During the analysis of the Vetscan OptiCell QCC, designated image analysis is employed to capture images of the fixed cell-like targets at various focal positions on the Quality Control Cartridge. Internal software extracts detailed features during the QCC analysis to report characteristics of the optical system performance. This provides additional verification on the optical module, illumination system, and mechanical components in order to verify the overall analyser performance.²

To perform an OptiCell QC Cartridge test, select a QC Test via the Vetscan Hub display. Under the 'My Lab' section, select the Vetscan OptiCell and select 'Create quality control test'. The reusable QC cartridge can be inserted into the OptiCell following the on-screen guidance. After the first scan, the OptiCell QC cartridge can be reused multiple times within a 90-day expiry window. The OptiCell QC cartridge barcode information will be read by the analyser and will alert the user once the 90-day expiry has been reached. When not in use, store the OptiCell QC cartridge in the original box packaging at ambient room temperature.

Reference: 2. Data on file. Study Report No. DHXMZ-US-24-235, 2024, Zoetis Inc.

4. Maintenance

4.2 Changing the Fan Filter

Periodically check the fan filter on the back panel of the Vetscan OptiCell analyser as needed to ensure proper ventilation and temperature control of the analyser. Check and replace the filter as needed if the analyser is in an environment with excessive dust or animal hair.*

***Note - Do not wash or reuse fan filters. Doing so may compromise the integrity of the fan filter material.**

Directions:

1. Turn the analyser off via the Vetscan Hub.
2. Remove the fan filter cover from the lower back panel.
3. Remove the fan filter and discard.
4. Place a new fan filter back into the cover.
5. Snap the filter cover back into place.
6. Power the analyser back on via the power button on the lower back panel.
7. The analyser will initialise and reconnect to Vetscan Hub.




Figure 9: Changing the Fan Filter

4. Maintenance

4.3 Software Updates

Periodic software updates and enhancements for Vetscan OptiCell will become available via Vetscan Hub. The software update procedure is automated, occurs during overnight hours and does not require any user interaction. The LED light band will turn blue during the software update process.*

***Note – Do not power off the Vetscan OptiCell analyser during the software update process.**

4.4 Service

- Vetscan OptiCell internal monitoring system monitors the health of the analyser and reports errors to the Vetscan Hub (see Troubleshooting Section). If the analyser does not work properly and needs to be serviced, please contact your local Zoetis Diagnostic Technical Support.
- Contact information can be found at <https://www.zoetisdiagnostics.com.au>.
- The warranty is void if the analyser is opened by a non-authorized service provider.

4.5 Disposal of the Vetscan OptiCell

For disposal of the used Vetscan OptiCell haematology cartridges, follow your local regulations. Do not dispose of the Vetscan OptiCell analyser as household waste. Contact Zoetis Diagnostic Technical Support for disposal.

5. Technical Support and Troubleshooting

5.1 Technical Support

Contact your local Zoetis Diagnostic Technical Support or distributor for technical assistance. Contact information can be found at <https://www.zoetisdiagnostics.com.au>

5.2 Errors and Warning Flags Displayed on Vetscan Hub

In certain cases, the OptiCell analyser will send warning flags to be displayed on the Vetscan Hub. Table 2 outlines the different warning messages for issues that may be encountered. Contact Zoetis Diagnostic Technical Support in case the corrective action does not solve the issue.

Table 2: Vetscan Hub errors and warning flags

Flag	Symbol on Affected Parameter	Interpretation
The blood sample may contain morphologically abnormal cells. Blood smear review recommended.	Asterisk (*)	Abnormal cells/cell morphology detected. Blood smear review recommended.
The blood sample may contain PLT clumps. Blood smear review recommended.	Asterisk (*)	The blood sample may contain PLT clumps. Blood smear review recommended. Consider sample quality and use proper sample handling techniques
Abnormal distribution of PLT cell volumes suspected. Blood smear review recommended.	Asterisk (*)	Abnormal distribution of PLT cell volumes detected. Blood smear review recommended.

5. Technical Support and Troubleshooting

5.3 Preanalytical Guidelines

1. Use K2/K3 EDTA anticoagulated whole blood sample tube, filled appropriately to the manufacturer's fill line.
2. Collected blood sample is at ambient room temperature of 17 °C to 27 °C.
3. Ensure minimum sample volume is met in the blood sample tube. Always fill to the manufacturer's fill line.
4. The EDTA sample tube is gently inverted 10-15 times immediately after sample collection, and again before starting the test.
5. The OptiCell Haematology Cartridge and the OptiCell Haematology Blood Sampler are valid (within the expiration date).
6. The OptiCell Haematology Blood Sampler capillary tubes are completely filled, without any air bubbles present.
7. The Vetscan OptiCell Haematology Blood Sampler is fully inserted into the Cartridge. Listen for a 'click' sound when fully inserted into the Vetscan OptiCell Haematology Cartridge.
8. Insert the prepared Vetscan OptiCell Haematology Cartridge immediately into the Vetscan OptiCell analyser.

5. Technical Support and Troubleshooting

5.4 Error Messages Displayed on Vetscan Hub

If a problem occurs during a test, Vetscan OptiCell analyser will send error messages and warning flags to the Vetscan Hub. For a list of error numbers and corresponding corrective actions, refer to Table 3. If the problem persists after the corrective action listed for the relevant error, contact Zoetis Diagnostic Technical Support.

Error Categories:

1XXXXX - System Errors

2XXXXX - Errors that are related to the OptiCell Haematology Cartridge or analyser during a measurement.

3XXXXX - Errors that are related to the patient blood sample.

4XXXXX - Errors that are related to operator error.

Table 3: Error Messages Displayed on Vetscan Hub

Error Number	Classification	Corrective Actions
110001 121001 123001 131117 122003 124001 131221	System Error	<p>System error occurred.</p> <ol style="list-style-type: none"> 1. Power off the analyser via Vetscan Hub and wait 30 seconds before powering back on. 2. Repeat the test using a new Vetscan OptiCell Haematology Cartridge and Vetscan OptiCell Haematology Blood Sampler. 3. If errors persist, contact Zoetis Diagnostic Technical Support.
136201	System Error	<p>System overheated.</p> <ol style="list-style-type: none"> 1. Power analyser off for 5 minutes via Vetscan Hub. Remove the fan filter and replace with a new filter. Then, restart the analyser. 2. If errors persist, contact Zoetis Diagnostic Technical Support.

5. Technical Support and Troubleshooting

Error Number	Classification	Corrective Actions
210011 231331 131331	Cartridge or System	<p>Focus/RBC error detected.</p> <ol style="list-style-type: none"> Repeat the test using a new Vetscan OptiCell Haematology Cartridge and Vetscan OptiCell Haematology Blood Sampler as follows: <ul style="list-style-type: none"> Verify both sampler capillaries are filled 100% to the end stopper. Verify that the sampler is inserted fully into the OptiCell Haematology Cartridge. Blood Sampler will 'click' upon full connection. Verify that there is no trapped air bubble or blood smear/debris on the OptiCell Haematology Cartridge chip area. If errors persist, contact Zoetis Diagnostic Technical Support.
231225 131225	Cartridge or System	<p>Chip Position Error detected.</p> <ol style="list-style-type: none"> Repeat the test using a new Vetscan OptiCell Haematology Cartridge and Vetscan OptiCell Haematology Blood Sampler. If the error persists, restart the analyser and repeat the test with another OptiCell Haematology Cartridge, preferably from a different lot. If a different lot is not available, perform an OptiCell Quality Control cartridge run. If errors persist, contact Zoetis Diagnostic Technical Support.
231332 131332	Cartridge	<p>Chip debris or bubbles detected.</p> <ol style="list-style-type: none"> Repeat the test with a new OptiCell Hematology Cartridge and verify there are no bubbles in the capillaries and no debris or smudges near the cartridge chip area. If the error persists, restart the analyzer. If possible, use a different lot and repeat the test using a new Vetscan OptiCell Hematology Cartridge and Vetscan OptiCell Hematology Blood Sampler. If a different lot is not available, perform an OptiCell Quality Control cartridge run. If errors persist, contact Zoetis Diagnostic Technical Support.

5. Technical Support and Troubleshooting

Error Number	Classification	Corrective Actions
		Cartridge vacuum error.
232221	System	<ol style="list-style-type: none"> 1.Repeat the test with a new OptiCell Haematology Cartridge. 2.If errors persist, contact Zoetis Diagnostic Technical Support.
		Invalid Scan Occurred
233222	Cartridge	<ol style="list-style-type: none"> 1.Repeat the test using a new Vetscan OptiCell Haematology Cartridge and Vetscan OptiCell Haematology Blood Sampler. 2.If errors persist, contact Zoetis Diagnostic Technical Support.
233223	Cartridge	Repeat the test using a new Vetscan OptiCell Haematology Cartridge and Vetscan OptiCell Haematology Blood Sampler. Verify there is no debris or smudges near the OptiCell Haematology Cartridge microfluidic chip area.
		Excessive Bubbles Detected
235658 235659 235655 135655 131445 231445	Cartridge or System	<ol style="list-style-type: none"> 1.Repeat the test using a new Vetscan OptiCell Haematology Cartridge and Vetscan OptiCell Haematology Blood Sampler. 2.Verify there are no air bubbles in the capillaries and no debris or smudges near the OptiCell Haematology Cartridge microfluidic chip area. 3.If the error persists, repeat the test with a new OptiCell Haematology Cartridge from another lot. If a different lot is not available, perform an OptiCell Quality Control cartridge run. 4.If errors persist, contact Zoetis Diagnostic Support.
		Preanalytical error suspected.
231443		<ol style="list-style-type: none"> 1.Repeat test using a new, fresh blood sample and confirm all preanalytical steps are correct. 2.If errors persist, contact Zoetis Diagnostic Technical Support.
		Differential Channel Error
231522	Blood Sample	<ol style="list-style-type: none"> 1.Repeat test 2.If the error persists, repeat the test with a new OptiCell Haematology Cartridge from another lot. If a different lot is not available, perform an OptiCell Quality Control cartridge run. 3.If errors persist, contact Zoetis Diagnostic Technical Support.

5. Technical Support and Troubleshooting

Error Number	Classification	Corrective Actions
235651	Blood Sample	<p>Preanalytical or Sample Quality error.</p> <ol style="list-style-type: none"> 1. Verify that both capillaries are filled 100% to the end stopper. 2. Verify that the OptiCell Haematology Blood Sampler is inserted fully 3. into the OptiCell Haematology Cartridge. Blood Sampler will 'click' upon full connection. Ensure proper sample mixing, and that no sample manipulation (i.e., dilution) was done to the blood sample. 4. Repeat the test using a fresh blood sample.
331333	Blood Sample	<p>Preanalytical error suspected.</p> <p>Repeat the test using a new Vetscan OptiCell Haematology Cartridge and Vetscan OptiCell Haematology Blood Sampler.</p> <ol style="list-style-type: none"> 1. Verify there are no bubbles in the sampler capillaries. 2. Verify that both capillaries are filled 100% to the end stopper. 3. Verify that the sampler is inserted fully into the cartridge and 'clicks' upon full connection.
335657	Blood Sample	<p>RBC per frame low.</p> <ol style="list-style-type: none"> 1. Potential low volume of sampled blood, or sample dilution occurred. 2. Repeat with a fresh blood sample and ensure proper sample handling techniques.

5. Technical Support and Troubleshooting

Error Number	Classification	Corrective Actions
335654 135655 131445 231445	Blood Sample	<p>Bubbles Detected.</p> <p>Repeat the test using a new Vetscan OptiCell Haematology Cartridge and Vetscan OptiCell Haematology Blood Sampler.</p> <ol style="list-style-type: none"> 1.Ensure proper sample handling techniques. 2.Verify that both capillaries are filled 100% to the end stopper and no bubbles are visible.
335656	Blood Sample	<p>Preanalytical error suspected.</p> <ol style="list-style-type: none"> 1.Repeat the test using a new Vetscan OptiCell Haematology Cartridge and Vetscan OptiCell Haematology Blood Sampler. 2.Ensure proper sample handling techniques. 3.Verify that both capillaries are filled 100% to the end stopper and no bubbles are visible.
335657	Blood Sample	<p>RBC ghost/haemolysis suspected.</p> <ol style="list-style-type: none"> 1.Repeat the test with a new OptiCell Haematology Cartridge and ensure proper sample handling techniques. 2.Verify that both capillaries are filled 100% to the end stopper and no bubbles are visible.

5. Technical Support and Troubleshooting

Error Number	Classification	Corrective Actions
440001	Operation error	Warning message: Used OptiCell Haematology Cartridge detected. Repeat test using a new OptiCell Haematology Cartridge.
440002	Operation error	Warning message: Expired OptiCell Haematology Cartridge detected. Repeat test using a new, not expired OptiCell Haematology Cartridge.
440003	Operation error	Warning message: Expired Quality Control cartridge detected. Repeat test with a new, not expired OptiCell QC cartridge.
440004	Cartridge Half Inside	Warning message: Cartridge inserted halfway. Confirm if cartridge was left inside of sample door. If so, remove the cartridge and press ok to continue.
440005	Cartridge Insert Timeout	Warning message: Timeout occurred. A 10-minute timeout error will occur after the test is started on the Vetscan Hub without inserting a cartridge into the analyser. Repeat the test and insert OptiCell Haematology Cartridge into the analyser immediately after starting the test.
440006	QC cartridge in Test mode	Warning message: A Quality Control cartridge was detected in sample test mode. Use an OptiCell Hematology Cartridge for patient sample testing.
440007	Haematology cartridge in QC mode	Warning message: OptiCell Hematology Cartridge was detected in QC Cartridge mode. Use an OptiCell QC cartridge for QC testing.

6. Performance Specifications

6.1 Vetscan OptiCell Reportable Ranges (Linearity)

Data was generated using available linearity controls and whole blood patient sample pools for RETIC.¹

Parameter	95% Lower Limit	95% Upper Limit	R ²
RBC (x 10 ⁶ cells/ μ L)	1.0	15.6	1.00
HGB (g/dL)	2.8	57.2	1.00
WBC (x 10 ³ cells/ μ L)	0.3	180	0.999
PLT (x 10 ³ cells/ μ L)	10	1900	0.999
RETIC (K/ μ L)	0	55	0.86

Reference: 1. TBD study on file.

7. Certifications

7.1 Countries of compliance

Countries of compliance available over Vetscan Hub.

7.2 USA Federal Communications Commission (FCC)

FCC CLASS B

- This equipment has been tested and found to comply with the limits for a Class B digital device, pursuant to Part 15 of the FCC Rules. These limits are designed to provide reasonable protection against harmful interference in a residential installation. This equipment generates, uses and can radiate radio frequency energy and, if not installed and used in accordance with the instructions, may cause harmful interference to radio communications. However, there is no guarantee that interference will not occur in a particular installation. If this equipment does cause harmful interference to radio or television reception, which can be determined by turning the equipment off and on, the user is encouraged to try to correct the interference by one or more of the following measures:
 - Reorient or relocate the receiving antenna.
 - Increase the separation between the equipment and receiver.
 - Connect the equipment into an outlet on a circuit different from that to which the receiver is connected.
 - Consult the dealer or an experienced radio/TV technician for help.

7.3 Industry Canada (IC)

WARNING This analyser complies with Industry Canada license-exempt RSS standard(s). Operation is subject to the following two conditions: (1) this device may not cause interference, and (2) this device must accept any interference, including interference that may cause undesired operation of the device.

7.4 Warranty

Contact your Zoetis representative for active warranty details, defect exclusions, and warranty coverage. If any defects in materials or workmanship occur in the analyser during active warranty period, Zoetis will repair or replace the analyser. The analyser's protection will be impaired if used in a manner not specified by the manufacturer. Service warranty to be rendered by Zoetis Diagnostic Technical Support. Contact information can be found at <https://www.zoetisdiagnostics.co.nz>.

8. Contact information

8.1 Manufacturer



Zoetis, Inc. 333 Portage
Street Kalamazoo, MI 49007,
USA www.zoetis.com

8.2 EU importer



Zoetis Belgium S.A. Rue Laid
Burniat 1, 1348 Louvain-La-Neuve,
Belgium

8.3 NZ importer



Zoetis New Zealand Pty Ltd.
Level 4, Mahuhu Crescent, Auckland, 1010

8.4 Technical support

0800 100 109
dxsupport.nz@zoetis.com

Learn more at
zoetisdiagnostics.co.nz

9. Symbols

Symbol	Definition/Use
	Manufacturer of the device
	Date of manufacturing at country of manufacturing
	Importer To indicate the identity and address of the Importer
	Model number
	Serial number
	Catalogue number
	Consult instructions for use or consult electronic instructions for use
	Caution Indicates that caution is necessary when operating the device or control close to where the symbol is placed, or that the current situation needs operator awareness or operator action in order to avoid undesirable consequences
	Electronic instructions for use. Scan QR code for access.
	Indoor use
	Direct current
	For veterinary use only
	WEEE Indicates that separate collection for waste of electric and electronic equipment (WEEE) is required
	Biological Risk Indicates that there are potential biological risks associated with the device

LOOK DEEPER

Manufacturer

Zoetis, Inc.
333 Portage Street
Kalamazoo, MI 49007, USA
www.zoetis.com

.....

EU Importer

Zoetis Belgium S.A.
Rue Laid Burniat 1,
1348 Louvain-La-Neuve, Belgium

.....

NZ Importer

Zoetis New Zealand Pty Ltd.
Level 4, Mahuhu Crescent,
Auckland, 1010

Learn more at zoetisdiagnostics.co.nz

Technical support

0800 100 109
dxsupport.nz@zoetis.com